

3707-711

No.VMGMC/Store/Q/ /2018
Office of the Dean,
Dr. V. M. Govt. Medical College,
Solapur Date:-09/03/2018

To


1. Health Diagnostic, Municipal Complex, 1st Floor, Main Road, Latur
2. Sainath Enterprises, 90, Majestic Shopping Centre, Saswad, Pune
3. Gunjana Surgical & Scientific Co., Mumbai
4. Rishi Surgical, Mumbai
- 5) N.P. Surgical Industries, Mumbai
- 6) Notice Board vmgmc, Solapur

Sub:- Quotation of Instruments.

Sealed quotations are invited for the following item from the stockiest for the use in various (Paed.) department of Dr. V. M. Govt. Medical College, Solapur.

Name of Item	Radiant Warmer With Baby Bassinet Machine Specifications
Radiant Warmer With Baby Bassinet Machine	<ol style="list-style-type: none"> 1. Should have facility to display both skin and air temperature separately. 2. Should have user friendly touch panel control 3. Should have ceramic infra-red heater. 4. Should have audiovisual alarm facility for overheating beyond set temperature range. 5. Should have alarm facility for patient temperature less than or greater than required temperature. 6. Should rotate and swivel in different directions, so as to allow taking X ray. 7. Light should be dazzle free. 8. Should have alarm for power failure, heater failure. 9. Should have alarm for probe failure. 10. Should have time out alarm in manual mode. 11. Should have built or provided along rechargeable battery to run equipment in case of power failure for at least ½ hour. 12. Should have facility to auto reset the system in case of hung up caused by power fluctuation. 13. Should have manual setting for high & low alarm setting. 14. In servo mode the heater output should be directly controlled to maintain the baby at required set temperature. 15. In manual mode heater output should be directly controlled by setting on the front panel. 16. Should have facility to lock the keyboard to avoid unwanted user. 17. The desired temperature range from 25 to 40degrees C. 18. Resolution should be 0.1 degrees C 19. The height of warmer should be adjustable. 20. Should have separate bassinet trolley. 21. Power input should be 220-240v AC, 50Hz fitted with Indian plug. 22. Shall meet IEC 0601-1-2001 OR FQULVALENT bis general requirement of safety for electromagnetic compatibility. 23. Comprehensive warranty for 5 year & 5 years of AMC after warranty. 24. User/technical manual to be supplied in English.

The sealed quotations may be sent in the name of the Dean, Dr. V. M. Govt. Medical College, Solapur on or before **due date 14/3/2018**. Quotations are received after due date will not be considered. Please subscribe the words on the top of envelop "Quotation for Instruments."
Quotation Opening Date 19/3/2018. Venue - College meeting hall.


DEAN,
DR. V. M. GOVT. MEDICAL COLLEGE,
SOLAPUR